



INSTRUCTIONS FOR USE –

AGAR PLATES



BAIRD PARKER AGAR

REF 1002

SPML manufacture and supply a range of Pre-prepared media for the culture of microorganisms such as bacteria yeasts and moulds. SPML offer a range of media pre-poured culture media to meet the needs of its customers across a range of industries that include Clinical, food, water, pharmaceutical, and many more.

The range of plates cover selective and non-selective media and plate sizes and are intended for the use by professional parties in the isolation of microbial organisms.

INTENDED USE :

Baird Parker Agar is a selective and diagnostic medium for the isolation and enumeration of Staphylococcus Aureus in foods.

• Micro Organism Reactions pH: 6.60 – 7.00

TECHNIQUE

Inoculate the medium with sample and Incubate the inverted dishes at 35°C. Examine after 24 hours and look for typical colonies of Staphylococcus aureus. Re-incubate negative cultures for a further 24 hours.

As methods and media preference may vary as to media types and test method due to jurisdictions, and personal preferences the customer should use the plates in accordance with their organizations stated methods and procedures however SPML product brochure does provide further product information .The end user should however take into consideration that selective media should, therefore, be compared with specimens/samples cultured on nonselective media to obtain additional information and help ensure recovery of potential pathogens and other significant organisms.

• Micro Organism Reactions pH: (6.60 – 7.00)

Organisms	Result
S. Epidermis	Good; black colony no halo
S. Aureus	Good; black colony with clear halo
P. Mirabilis	Good; brown/black colony
E.Coli	No Growth

References.

¹ EN 12322:1999 - In vitro diagnostic medical devices - Culture media for microbiology - Performance criteria for culture media.

² Clinical and Laboratory Standards Institute. 2004. Approved standard: M22-A3, Quality control for commercially prepared microbiological culture media, 3rd ed

³ ISO11133 Microbiology of food, animal feed and water — Preparation, production storage and performance testing of culture media



Medical Device & QA Services Ltd (MDQAS),

Spring Court, Spring Road,
HALE, Cheshire. WA14 2UQ.
United Kingdom.

Tel: +44 (0) 845 527 5078, Fax: +44 161 903 9787
Email: info@mdqas.com Web
Site: www.mdqas.com

PRECAUTIONS



. For professional use only.

Do not use plates if they show evidence of microbial contamination, discoloration, drying, cracking or other signs of deterioration.

Consult In house instructions to ensure correct application of product use is observed. Ensure GLP and aseptic handling procedures are followed, used plates should be treated as biohazards, and disposal of used product should be treated as such, and disposed of in accordance with local /national regulations.

STORAGE AND SHELF LIFE

On receipt, store plates in the dark at 2 to 8° C, in their original sleeve wrapping until just prior to use. Avoid freezing and overheating. The plates may be inoculated up to the expiration date (see package label) and incubated for the recommended incubation times.

Plates from opened stacks of 10 plates can be used for one week when stored in a clean area at 2 to 8° C.

QUALITY CONTROL

SPML undertake batch release for all media checking pH, sterility, Inhibition and enhancement testing of all media batches based on the EN12322:1999¹, M22-A3², and ISO11133³,

For clinical purpose the product is CE marked in accordance with the requirements of 97/79/EC IVD Directive and registered with the Competent Authority MHRA in the United Kingdom. SPML is ISO9001:2008 certified

FURTHER INFORMATION

For further information please contact your local SPML representative.

Saudi Prepared Media Laboratory Company Ltd

Telephone: +966 11 4767931, 11 4773697

FAX: +966 1 4778313

E-mail: media@spml.com.sa

Web Site: <http://www.spml.com.sa>



Saudi Prepared Media Laboratory Company Ltd

Al Harath Bin Al Habab Street ٨٩٦٨

High-al Dubbath " Malaz"

Riyadh 11461

Kingdom Of Saudi Arabia



SPMLSELECT

INSTRUCTIONS FOR USE –

AGAR PLATES



BAIRD PARKER AGAR

REF 1002

Symbols Definition

	Product Reference Number
	For in vitro diagnostic use
	Do not use if damaged
	Sterile by Moist / Dry heat
	Non Sterile
	Single Use
	Manufacturer
	Lot Number
	Bio hazardous material
	CE Mark – Against the European In Vitro Diagnostic Medical Device Directive 98/79/EC
	Storage temperature limitation
	Manufacture date
	Consult Instructions for use
	Avoid direct sunlight
	Keep Dry
	Expiration Date
	EU Representative
	WARNING – Consult IFU